


# PRODUKTDATENBLATT



<b>Produkt</b>	Arvikon Arztkittel steril einweg
<b>Artikelnr.</b>	CAP-med-s-023
<b>Verpackungseinheit (VPE)</b>	1/1 einzeln verpackt oder 60 Stk. im Karton (Verpackung in Englisch)
<b>Material</b>	45g, Spunbond + Meltblown + Spunbond Nonwovens (SMS) Material
<b>Farbe / Größen</b>	blau oder grün – M / L / XL
<b>Zertifizierung, Herstellung</b>	CE (TÜV SÜD) Notified Body Identification-Nr. 0123, CFDA, ISO 13485 und EN 13795 zertifiziert
<b>Schutzklasse</b>	Schutzklasse 3, 99,9% Hohe Bakterienfiltereffizienz
<b>Anwendungsbereich</b>	Sämtliche Anwendungen im Stations- und OP-Bereich
<b>Abmessungen / Gewicht</b>	50x40x45 in cm (HxBxT), 7,2kg pro Karton

Größen	M	L	XL	Toleranz
Abmessungen / cm	145	150	150	+/-2 cm

CAPULUS übernimmt keine Gewährleistung/ Haftung für die Aktualität, Richtigkeit und Vollständigkeit der vom Hersteller bereitgestellten Informationen und Daten. Das Produkt kann vom o.g. Bild abweichen.

**Kontakt / Bestellungen:**

Tel: +49 40 307 772 15  
 E-Mail: info@capulus-services.de  
 Anfragen: Bestellformular Geschäftskunden  
 Internet: CAPULUS Medical Services

# ZERTIFIZIERUNG

TÜV SÜD  
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認許證書 ◆ CERTIFICADO ◆ CERTIFICAT



Deutsche  
Akkreditierungsstelle  
D-ZM-11321-01-00





Product Service

## Certificate

No. Q6 083528 0010 Rev. 00

**Holder of Certificate:** [Redacted]

**Facility(ies):** [Redacted]

**Certification Mark:**



tuv-sud.com/ps-cert

**Scope of Certificate:** Production and Distribution of Surgical Set, Surgical Gown, Surgical Drape, Tube Cover, Liquid Collection Pouch, Protective Gown, Warm Blanket and Surgical Hood

**Applied Standard(s):** EN ISO 13485:2016  
Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)  
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system (excluding subclause 7.3), which meets the requirements of the listed standard(s). See also notes overleaf.

**Report No.:** SH1978307

**Valid from:** 2019-04-01  
**Valid until:** 2022-03-31

**Date,** 2019-04-01

  
 Stefan Preiß

Page 1 of 1  
TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

